Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D0028568	(X3) Date Survey Completed 07/24/2019
Name of Provider or Supplier Family Medical Clinic Of Harrogate, Pc	**Street Address, City, State** 170 Beech St, Suite 1, Harrogate, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6053	TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9) The technical consultant is responsible for evaluating and documenting the performance of individuals responsible for moderate complexity testing at least semiannually during the first year the individual tests patient specimens. This STANDARD is not met as evidenced by: ========== Based on a review of the laboratory's employee competency evaluations and upon interview with the current Technical Consultand and Testing Personnel, determined the Technical Consultant failed to perform and document a semi-annual competency evaluation for testing personnel #3 who began laboratory testing 9/7/2017. The findings include: 1. A review of the laboratory's semi-annual competency evaluation results were not available for testing person #3 of 3 who began patient testing for Complete Blood Counts (CBC) in September of 2017. 2. An interview with the current Technical Consultant and Testing Personnel at approximately 12:30 pm on July 24, 2019 confirmed that no semi-annual competency evaluation was documented for testing person #3 since training date 9/7/2017. ==========