

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  44D0234042	<b>(X3) Date Survey Completed</b>  08/26/2025
<b>Name of Provider or Supplier</b>  Tennessee Urology Associates Pllc	<b>Street Address, City, State</b>  800 Oak Ridge Turnpike, Suite A-101, Oak Ridge, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5217</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by:</p> <ul style="list-style-type: none"> <li>. Based on laboratory observation, a review of patient test reports, laboratory records, and interviews with the laboratory director and manager, the laboratory failed to verify the accuracy of histopathology twice annually in 2023, 2024, and through the survey date in 2025 (08.26.2025). The findings include: 1. Observation of the laboratory on 08.26.2025 at 10:10 a.m. revealed equipment used for processing and staining tissue for slide interpretation for histopathology procedures. Subsequent observation of the laboratory director's office on 08.26.2025 at 11:45 a.m. revealed a microscope for slide interpretation for histopathology. 2. Review of patient test reports revealed the following histopathology slide interpretations: -Patient PR23-00936 on 12.04.2023 -Patient PR24-01223 on 11.18.2024 -Patient PR25-00250 on 03.11.2025 3. A review of the laboratory's records revealed that it did not verify the accuracy of histopathology slide interpretation twice annually in 2023 and 2024 and had not performed it in 2025 through the survey date, 08.26.2025. 4. Interviews with the laboratory manager on 08.26.2025 at 11:10 a.m. and with the laboratory director on 08.26.2025 at 11:50 a.m. confirmed the above survey findings.</li> </ul>