

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D0309288	(X3) Date Survey Completed 10/31/2019
Name of Provider or Supplier Paul G Smith Jr Do Plc	Street Address, City, State 2121 North Ocoee Street Suite 101, Cleveland, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: ===== Based on review of Proficiency Testing (PT) scores from 2018 and 2019 and interview with lead laboratory testing personnel, the laboratory failed to prove twice a year accuracy verification for microscopic urine sediment. The findings include: 1. A review of PT scores for 2018 and 2019 showed unacceptable scores of 50% for one of two events in 2018 and 2019. 2. An interview with the lead laboratory testing personnel at approximately 12:45 on October 31, 2019 confirmed the laboratory failed to prove twice a year accuracy verification through PT for urine sediment during 2018 and 2019. =====</p>