

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D0312039	(X3) Date Survey Completed 07/22/2024
Name of Provider or Supplier Scott County Community Hospital, Inc	Street Address, City, State 18797 Alberta St, Oneida, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D3011	<p>FACILITIES CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by: Intakes: TN00070425 Based on observation of the laboratory area and staff interview, the facility failed to ensure the safety and protection of laboratory reagents and personnel from potential exposure to biohazardous waste material that leaked into the laboratory walk-in refrigerator on July 15, 2024. The findings include: 1. Observation of the laboratory area on 07.22.2024 at 9:28 a.m. revealed a wood-lined walk-in refrigerator for storing the laboratory's reagents. Also observed were noticeable stains and evidence of residue in the middle of the refrigerator located on the ceiling and the floor directly below the stains on the ceiling. 2. During an interview on 07.22.2024 at 9:30 a.m., the Laboratory Supervisor stated, "The facility had more plumbing issues, and the waste leaked into the walk-in refrigerator on July 15th." She also confirmed that the stains and remaining residue on the ceiling and floor of the walk-in refrigerator were evidence of biohazardous wastewater from a busted sewer pipe that leaked into the refrigerator on 07.15.2024.</p>