Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	44D0312905	02/11/2025
Name of Provider or Supplier	**Street Address, City, State**	
Anne Arundel Dermatology, Pa	123 Fox Rd, Knoxville, TN	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5401	PROCEDURE MANUAL CFR(s): 493.1251(a) (a) A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens. This STANDARD is not met as evidenced by: Intakes: TN00073355 Based on observation of the laboratory, a review of the laboratory procedure manual, and an interview with the histology laboratory supervisor, it was determined that the laboratory failed to follow the Standard Operating Procedure Manual for labeling chemical reagents. The findings include: 1. Observation of the laboratory on 02.11.2025 at 09:13 a.m. revealed a flammable cabinet with three containers of unlabeled substance, verbally identified as Xylene waste by histology laboratory supervisor. 2. A review of the laboratory's Standard Operating Procedure Manual under " Storage, use and handling" revealed the following: "All reagents are stored according to manufacturer's guidelines along with all state and federal Guidelines and are kept in a flammable cabinet. If chemicals are expired they are not used and disposed of. All chemicals are labeled clearly on outside of all bottles. Material safety data sheets are in the back of binder for reference." 3. An interview with the histology laboratory supervisor on 02.11.2025 at 09:13 a.m. confirmed the findings.