

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D0312905	(X3) Date Survey Completed 02/13/2025
Name of Provider or Supplier Anne Arundel Dermatology, Pa	Street Address, City, State 123 Fox Rd, Knoxville, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5415	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(c)</p> <p>(c) Reagents, solutions, culture media, control materials, calibration materials, and other supplies, as appropriate, must be labeled to indicate the following: (c)(1) Identity and when significant, titer, strength or concentration. (c)(2) Storage requirements. (c)(3) Preparation and expiration dates. (c)(4) Other pertinent information required for proper use.</p> <p>This STANDARD is not met as evidenced by: Based on observation of laboratory and staff interviews, the laboratory failed to label reagents used in the Histopathology tissue processing laboratory with information when transferred from the primary to the secondary container. The findings include: 1. Observation of the Histology laboratory on 02.11.2025 at 9:00 a.m. revealed a secondary container used for processing samples for Histopathology procedures that was not labeled with identity, lot number, and expiration date when transferred from the original container to the secondary container. The reagent was verbally identified as Nail Prep (Item #NP-500) by histology laboratory supervisor. 2. An interview with the histology laboratory supervisor on 02.11.2025 at 09:00 a.m. confirmed the survey findings.</p>
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by:</p>

Citation One: Based on observation of the Histology laboratory and staff interview, the laboratory failed to ensure that tissue marking dyes and hematoxylin used for inking and staining had not expired on the date of the survey (02.11.2025). The findings include: 1. Observation of the Histology laboratory on 02.11.2025 at 9:00 am revealed the following expired dyes and stains in use. -Mercedes Scientific Tissue Marking Dye Red (2 bottles), lot number 23004, expiration date 01.31.2025 - Modified Mayers Hematoxylin (1 bottle), lot number 181170, expiration date 10.31.2024. 2. An interview with the Histology laboratory supervisor on 02.11.2025 at 09:03 a.m. confirmed the above findings. Citation Two: Based on observation of the MOHS processing room and staff interview, the laboratory failed to ensure it did not use tissue marking dyes for MOHS Histopathology tissue processing past their expiration date on the survey date (02.11.2025). The finding include: 1. Observation of the MOHS processing room on 02.11.2025 at 9:39 a.m. revealed the following tissue marking dyes in use. -Cancer Diagnostic, Inc. Tissue Marking Dyes (2 bottles) Orange, lot number 21111, expiration date 04.30.2023. Green, lot number 21130, expiration date 05.31.2023. 2. Interview with the MOHS technician on 02.11.2025 at 9:39 a.m. confirmed the above findings. Work Key: MOHS: micrographically oriented histography surgery