

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D0857774	(X3) Date Survey Completed 11/20/2025
Name of Provider or Supplier Pediatric Associates Of Davidson County	Street Address, City, State 2201 Murphy Avenue Suite 201, Nashville, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6018	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(iii)</p> <p>(e)(4)(iii) All proficiency testing reports received are reviewed by the appropriate staff to evaluate the laboratorys performance and to identify any problems that require corrective action; and</p> <p>This STANDARD is not met as evidenced by: Based on a review of the laboratory's personnel job descriptions, College of American Pathologists (CAP) proficiency testing (PT) records, lack of documentation, and staff interviews, the laboratory director failed to identify PT results that required corrective action for the complete blood count with automated differential (CBC w/Diff) analytes for one of five PT events reviewed in 2024 and 2025. The findings include:</p> <ol style="list-style-type: none"> 1. A review of the laboratory director's job description in the section, "Example of Duties," revealed the following: "Ensure all proficiency-testing reports received are reviewed by the appropriate staff to evaluate the laboratory's performance and to identify any problems that require corrective action". 2. A review of the laboratory's 2024 Event Three CAP PT performance evaluation for the CBC w/Diff analytes, reviewed by the laboratory director on 11/13/2024, revealed that the platelet analyte for sample FH1-13 was unacceptable. (Reported Result 54/Acceptable Range 58-98) 3. Documentation of corrective actions for sample FH1-13 for the 2024 Event Three was unavailable on the date of the survey (11/20/2025). 4. An interview with the laboratory director and office manager on 11/20/2025 at 1:00 p.m. confirmed the survey findings.