

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D0864535	(X3) Date Survey Completed 04/09/2025
Name of Provider or Supplier Tennessee Oncology Pllc	Street Address, City, State 2004 Hayes Street, Suite 350, Nashville, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5775	<p>COMPARISON OF TEST RESULTS CFR(s): 493.1281(a)(c)</p> <p>(a) If a laboratory performs the same test using different methodologies or instruments, or performs the same test at multiple testing sites, the laboratory must have a system that twice a year evaluates and defines the relationship between test results using the different methodologies, instruments, or testing sites.</p> <p>This STANDARD is not met as evidenced by: Based on laboratory observation, review of the laboratory procedure, instrument correlation records, and staff interviews, the laboratory failed to evaluate the relationship between test results between instruments performing complete blood count (CBC) testing twice per year in 2024 and 2025. The findings include: 1. Observation of the laboratory on 04/09/2025 at 8:30 a.m. revealed one Sysmex XN-550 (ID: 12532) and one Sysmex XN-430 (ID: 12034) performing patient CBC testing. 2. A review of the laboratory's "Intra-Instrument Comparison" procedure revealed the requirement that the laboratory must correlate patient results between hematology analyzers at least semiannually. 3. A review of the laboratory's 2024 and 2025 instrument comparison records for the hematology analyzers revealed that the laboratory performed an instrument comparison on 04/25/2024. At the time of the survey (04/09/2025), there was no documentation of a second comparison between the Sysmex XN-550 and Sysmex XN-430. 4. An interview on 04/09/2025 at 11:30 a.m. with the technical consultants confirmed the laboratory failed to correlate patient results between their hematology analyzers twice in 2024.</p>