Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D0864535	(X3) Date Survey Completed 04/17/2026
Name of Provider or Supplier Tennessee Oncology Pllc	**Street Address, City, State** 2004 Hayes Street, Suite 350, Nashville, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d) (d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality. This STANDARD is not met as evidenced by: Based on observation of the laboratory and staff interviews, the laboratory failed to ensure that capillary ethylenediaminetetraacetic acid (EDTA) blood collection tubes were not used past the manufacturer's expiration date for complete blood count with differential (CBC w/Diff) patient testing in 2025. The findings include: 1. Observation of the laboratory on 04/17/2026 at 9:45 am revealed the Sysmex XN 430 (Serial 12034) and Sysmex XN 530 (Serial 12532) used to perform CBC w/Diff patient testing with EDTA venipuncture and capillary blood collection tubes. An observation of the adjoining phlebotomy room stations revealed a drawer that contained mixed lots of EDTA capillary blood collection tubes that included lot number 23C4349 with a manufacturer's expiration date of 03/31/2025. 2. An interview with technical consultants one and two on 04/17/2026 at 11:45 a.m. confirmed the survey findings.