

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  44D0892367	<b>(X3) Date Survey Completed</b>  11/03/2021
<b>Name of Provider or Supplier</b>  Lebonheur Pediatrics, Llc	<b>Street Address, City, State</b>  3876 New Covington Pike, Memphis, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5215</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2)</p> <p>The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results).</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory records and staff interviews, the laboratory failed to evaluate the accuracy of non-graded proficiency testing results in 2019 and 2020 (two of two non-graded scores). The findings include: 1. Review of the laboratory's proficiency testing performance evaluation reports revealed the following: 2019 event one--Sample number US-02 scored as "Not Graded" with no evaluation of the results to determine the laboratory's accuracy. 2020 event three--Sample number US-06 scored as "Not Graded" with no evaluation of the results to determine the laboratory's accuracy. 2. Interview with the technical consultant on 11/03/2021 at approximately 4pm confirmed the laboratory failed to evaluate the accuracy of non-graded proficiency testing results in 2019 and 2020 (two of two non-graded proficiency testing scores).</p>