

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D0921688	(X3) Date Survey Completed 02/13/2026
Name of Provider or Supplier Jackson Clinic,Pa North Internal, The	Street Address, City, State 2863 Hwy 45 Bypass, Jackson, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The following deficiencies are a result of a desk review of proficiency testing scores obtained from the national database and verified with the proficiency testing company. The facility was found to be out of compliance with the conditions of the CLIA program. The following CONDITION LEVEL DEFICIENCIES were found to be out of compliance: D2016 - 42 C.F.R. 493.803 Condition: Successful participation [proficiency testing] D6076 - 42 C.F.R. 493.1441 Condition: Laboratories performing high complexity testing; laboratory director
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on a desk review of proficiency testing (PT) records from the Certification and</p>

	<p>Survey Provider Enhanced Reporting (CASPER) 0155 report and College of American Pathologists (CAP) proficiency testing records, the laboratory failed to successfully participate in a proficiency testing program approved by HHS, for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. The laboratory failed to successfully participate in the subspecialty of Bacteriology. Refer to 2028.</p>
<p>D2028</p>	<p>BACTERIOLOGY CFR(s): 493.823(e)</p> <p>(e) Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on a proficiency testing desk review of the CASPER 0155 report and CAP 2025 proficiency testing records, the laboratory failed to achieve an overall satisfactory performance (80% or greater) for two of three consecutive testing events in the subspecialty of Bacteriology. The findings included: 1. A review of the CASPER 0155 report revealed the following: Bacteriology 2025-1st Event The laboratory received an unsatisfactory score of 78% for Bacteriology. Bacteriology 2025-3rd Event The laboratory received an unsatisfactory score of 54% for Bacteriology. 2. A review of the laboratory's CAP 2025 proficiency testing records confirmed the laboratory received the above results. The laboratory received an overall score of 78% in Event One and 54% in Event Three for Bacteriology.</p>
<p>D6076</p>	<p>LABORATORY DIRECTOR CFR(s): 493.1441</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1443 of this subpart and provides overall management and direction in accordance with 493.1445 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on a proficiency testing desk review of CASPER 0155 Report and CAP 2025 proficiency testing records, the laboratory director failed to provide overall management and direction of the laboratory services. The laboratory director failed to ensure proficiency testing samples were tested as required. Refer to D6089.</p>
<p>D6089</p>	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(4)(i)</p> <p>(e)(4)(i) The proficiency testing samples are tested as required under subpart H of this part;</p> <p>This STANDARD is not met as evidenced by: Based on a proficiency testing desk review of CASPER 0155 report and CAP 2025 proficiency testing records, the laboratory director failed to ensure proficiency testing samples were tested as required. The laboratory director failed to ensure successful participation in an HHS approved proficiency testing program. Refer to D2028.</p>