

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D0933099	(X3) Date Survey Completed 06/03/2021
Name of Provider or Supplier Thomas P Chu Md Pc	Street Address, City, State 520 Trinity Creek Cove, Cordova, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D3031	<p>RETENTION REQUIREMENTS CFR(s): 493.1105(a)(3)</p> <p>Analytic systems records. Retain quality control and patient test records (including instrument printouts, if applicable) and records documenting all analytic systems activities specified in 493.1252 through 493.1289 for at least 2 years.</p> <p>This STANDARD is not met as evidenced by: Based on observation of the laboratory, review of patient test reports, request for laboratory records and interview with the laboratory director, the laboratory failed to retain records of lot numbers for reagents used for performing histopathology tests in 2019, 2020, and 2021. The findings include: 1. Observation of the laboratory on June 3, 2021 at 10am revealed a microscope in use for reading slides for histopathology, as well as inks and reagents used for inking, processing and staining of tissue removed during MOHS procedure. 2. Review of patient test reports from 2019, 2020, and 2021 (patient numbers one through five) revealed testing performed for histopathology on tissue removed during MOHS procedure. 3. Request for records of reagent lot numbers, expiration dates, and dates in use revealed no records were maintained. 4. Interview with the laboratory director on June 3, 2021 at 11:00 am confirmed the laboratory does not maintain records of lot numbers, expiration dates and dates in use for reagents, stains and inks used in performing histopathology procedures.</p>
D6168	<p>TESTING PERSONNEL CFR(s): 493.1487</p> <p>The laboratory has a sufficient number of individuals who meet the qualification requirements of 493.1489 of this subpart to perform the functions specified in 493.1495 of this subpart for the volume and complexity of testing performed.</p>

This CONDITION is not met as evidenced by:
The laboratory personnel performing high complexity testing did not meet the regulatory requirements for high complexity testing personnel. (Refer to D6171)

D6171

TESTING PERSONNEL QUALIFICATIONS
CFR(s): 493.1489(b)

(b) Meet one of the following requirements: (b)(1) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located or have earned a doctoral, master's or bachelor's degree in a chemical, physical, biological or clinical laboratory science, or medical technology from an accredited institution; (b)(2)(i) Have earned an associate degree in a laboratory science, or medical laboratory technology from an accredited institution or-- (b)(2)(ii) Have education and training equivalent to that specified in paragraph (b)(2)(i) of this section that includes-- (b)(2)(ii)(A) At least 60 semester hours, or equivalent, from an accredited institution that, at a minimum, include either-- (b)(2)(ii)(A)(1) 24 semester hours of medical laboratory technology courses; or (b)(2)(ii)(A)(2) 24 semester hours of science courses that include-- (b)(2)(ii)(A)(2)(i) Six semester hours of chemistry; (b)(2)(ii)(A)(2)(ii) Six semester hours of biology; and (b)(2)(ii)(A)(2)(iii) Twelve semester hours of chemistry, biology, or medical laboratory technology in any combination; and (b)(2)(ii)(B) Have laboratory training that includes either of the following: (b)(2)(ii)(B)(1) Completion of a clinical laboratory training program approved or accredited by the ABHES, the CAHEA, or other organization approved by HHS. (This training may be included in the 60 semester hours listed in paragraph (b)(2)(ii)(A) of this section.) (b)(2)(ii)(B)(2) At least 3 months documented laboratory training in each specialty in which the individual performs high complexity testing. (b)(3) Have previously qualified or could have qualified as a technologist under 493.1491 on or before February 28, 1992; (b)(4) On or before April 24, 1995 be a high school graduate or equivalent and have either-- (b)(4)(i) Graduated from a medical laboratory or clinical laboratory training program approved or accredited by ABHES, CAHEA, or other organization approved by HHS; or (b)(4)(ii) Successfully completed an official U.S. military medical laboratory procedures training course of at least 50 weeks duration and have held the military enlisted occupational specialty of Medical Laboratory Specialist (Laboratory Technician); (b)(5)(i) Until September 1, 1997-- (b)(5)(i)(A) Have earned a high school diploma or equivalent; and (b)(5)(i)(B) Have documentation of training appropriate for the testing performed before analyzing patient specimens. Such training must ensure that the individual has-- (b)(5)(i)(B)(1) The skills required for proper specimen collection, including patient preparation, if applicable, labeling, handling, preservation or fixation, processing or preparation, transportation and storage of specimens; (b)(5)(i)(B)(2) The skills required for implementing all standard laboratory procedures; (b)(5)(i)(B)(3) The skills required for performing each test method and for proper instrument use; (b)(5)(i)(B)(4) The skills required for performing preventive maintenance, troubleshooting, and calibration procedures related to each test performed; (b)(5)(i)(B)(5) A working knowledge of reagent stability and storage; (b)(5)(i)(B)(6) The skills required to implement the quality control policies and procedures of the laboratory; (b)(5)(i)(B)(7) An awareness of the factors that influence test results; and (b)(5)(i)(B)(8) The skills required to assess and verify the validity of patient test results through the evaluation of quality control values before reporting patient test results; and (b)(5)(i)(B)(8)(ii) As of September 1, 1997, be qualified under 493.1489(b)(1), (b)(2), or (b)(4), except for those individuals qualified under paragraph (b)(5)(i) of this section who were performing high complexity testing on or before April 24, 1995; (b)(6) For blood gas analysis-- (b)(6)

(i) Be qualified under 493.1489(b)(1), (b)(2), (b)(3), (b)(4), or (b)(5); (b)(6)(ii) Have earned a bachelor's degree in respiratory therapy or cardiovascular technology from an accredited institution; or (b)(6)(iii) Have earned an associate degree related to pulmonary function from an accredited institution; or (b)(7) For histopathology, meet the qualifications of 493.1449 (b) or (l) to perform tissue examinations.

This STANDARD is not met as evidenced by:

Based on observation of the laboratory, interview with testing personnel number two, request for personnel records, and interview with the office manager, the personnel performing high complexity testing failed to meet the regulatory education and experience requirements. The findings include: 1. Observation of the laboratory on June 3, 2021 at 10:00am revealed a microscope, reagents, stains and inks in use for histopathology testing procedures. 2. Interview with testing personnel number two on June 3, 2021 at 10:00am revealed that testing personnel number two and testing personnel number three apply marking ink to tissue removed during MOHS procedure. 3. Request for personnel records for testing personnel number two and three for highest level of education revealed no records were available. 4. Interview with the office manager on June 3, 2021 at 11:30am confirmed no personnel records were available for testing personnel number two and testing personnel number three. Testing personnel number two and three perform high complexity testing to include inking of tissue removed during MOHS procedures.