

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D0934695	(X3) Date Survey Completed 01/09/2020
Name of Provider or Supplier Premier Pediatrics	Street Address, City, State 101 Med Tech Parkway, Suite 405, Johnson City, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6054	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9)</p> <p>The technical consultant is responsible for evaluating and documenting the performance of individuals responsible for moderate complexity testing at least annually, after the first year.</p> <p>This STANDARD is not met as evidenced by:</p> <p>===== Based on a review of the laboratory's employee competency evaluations and upon interview with the Laboratory Coordinator, determined the Technical Consultant failed to perform and document annual competency evaluations for 2 of 3 testing personnel for 2019. The findings include: 1. A review of employee competency evaluations for testing personnel #1 and #3 disclosed no annual competency evaluations documented for 2019. 2. Interview with Laboratory Coordinator at approximately 12:15 p.m. on January 9, 2020 confirmed the Technical Consultant failed to perform and document annual competency evaluations for testing personnel #1 and #3 for 2019.</p> <p>=====</p>