

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D0955426	(X3) Date Survey Completed 08/14/2018
Name of Provider or Supplier Lebonheur Pediatrics Llc	Street Address, City, State 1255 S Germantown Rd, Germantown, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2127	<p>HEMATOLOGY CFR(s): 493.851(d)</p> <p>Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of 0 for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on observation of the laboratory, review of the 2017 proficiency testing (PT) records and interview with the technical consultant, the laboratory failed to participate in the 2017 event two for the automated three part white blood cell (wbc) differential (diff). The findings include: 1) Observation of the laboratory on August 15, 2018 at 1:40 p.m. revealed the Beckman Counter AcT diff instrument (system ID 6417399) in use for patient CBC testing, which includes automated three part wbc diff. 2) Review of the 2017 event two PT records revealed zero percent (%) score for no participation for the wbc diff. 3) Interview on August 15, 2018 at 2:45 p.m. with the technical consultant confirmed the laboratory received a zero % score for the automated three part wbc diff in the PT 2017 event two.</p>