

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D0963940	(X3) Date Survey Completed 10/01/2025
Name of Provider or Supplier Syed A A Zaidi Md Pllc	Street Address, City, State 868 Highway 51 South, Ripley, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on laboratory observation and staff interview, the laboratory failed to ensure that tubes used for collecting fingerstick Complete Blood Count with Automated White Blood Cell differential were not used past the expiration date of 05/31/25. The findings include: 1. Laboratory observation on 10/01/25 at 9:15 a.m. revealed the Sysmex XN 330 used for performing patient testing for CBC w/Diff. The phlebotomy area had a drawer with tubes used for collecting fingerstick CBC w/Diff samples. The observed lot number was 23E4127 with an expiration date of 05/31/25. 2. The technical consultant confirmed the survey findings during an interview on 10/01/25 at 2:30 p.m.</p>