

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  44D0978864	<b>(X3) Date Survey Completed</b>  05/17/2018
<b>Name of Provider or Supplier</b>  Cherokee Health Systems	<b>Street Address, City, State</b>  6350 West Andrew Johnson Hwy, Talbott, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2016</b>	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: _____ The laboratory failed to maintain satisfactory participation in two out of two events for the automated White Blood Cell (WBC) Differential, resulting in the first unsuccessful Proficiency Testing (PT) occurrence for the automated WBC Differential analyte (Refer to D2130). _____</p>
<b>D2130</b>	<p>HEMATOLOGY CFR(s): 493.851(f)</p>

Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

\_\_\_\_\_ Based on a desk review of the CMS CASPER Report 0155D and the Laboratory's 2017 and 2018 Proficiency Testing (PT) Program results, determined the Laboratory failed to maintain satisfactory performance in two consecutive events for the WBC Auto Differential, event three of 2017 and event one of 2018, resulting in the first unsuccessful occurrence. The findings include: 1. A review of the CMS CASPER Report 0155D for Proficiency Testing revealed the WBC Auto Diff scores as 4% for event three of 2017 and 0% for event one of 2018. 2. A review of the PT Program results for the WBC Auto Differential revealed an unsatisfactory score as 4% for event three of 2017 and 0% score for event one of 2018, resulting in the first unsuccessful occurrence for the WBC Auto Differential. \_\_\_\_\_