

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D0989207	(X3) Date Survey Completed 06/25/2024
Name of Provider or Supplier Affiliated Dermatologists Green Hills Pc	Street Address, City, State 4301 Hillsboro Road Suite 200, Nashville, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the laboratory's quality assurance (QA) policy, lack of documentation, and staff interviews, the laboratory failed to verify the accuracy of micrographically oriented histographic surgery (MOHS) testing at least twice annually in 2022 and 2023. The findings include: 1. A review of the policy titled "Quality Assurance Manual" revealed the following statement: - "Any test performed in the laboratory for which proficiency testing is not available will be verified at least twice a year, and the results will be reviewed by the Laboratory Director." 2. No verification of accuracy documents were available for MOHS testing in 2022 and 2023. 3. An interview with the MOHS histotechnician on 06/25/2024 at 10:30 a.m. confirmed that the laboratory did not verify the accuracy of MOHS testing in either 2022 or 2023.</p>