

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 44D0993884	<b>(X3) Date Survey Completed</b> 10/05/2021
<b>Name of Provider or Supplier</b> Proactive Heart And Vascular, Pllc	<b>Street Address, City, State</b> 7751 Wolf River Blvd, Germantown, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D6052</b>	<p><b>TECHNICAL CONSULTANT RESPONSIBILITIES</b> CFR(s): 493.1413(b)(8)(vi)</p> <p>The procedures for evaluation of the competency of the staff must include, but are not limited to assessment of problem solving skills.</p> <p>This STANDARD is not met as evidenced by: Based on review of competency assessment documents and interview with the laboratory liaison, the technical consultant failed to assess problem solving skills for the sole testing person in 2020 and 2021. 1. Review of competency assessment documents revealed that assessment of problem solving skills was not included in the annual competency documentation for the sole testing person in 2020 and 2021. 2. Interview with the laboratory liaison on October 5, 2021 at 10:45 am confirmed the technical consultant failed to include assessment of problem solving when performing annual competency for the sole testing person in 2020 and 2021.</p>