

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D1013778	(X3) Date Survey Completed 01/15/2026
Name of Provider or Supplier Cumberland Dermatology, Pc	Street Address, City, State 29 Taylor Ave, Suite 101, Crossville, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation of the laboratory and staff interviews, the laboratory failed to ensure that it did not use five of six lots of tissue marking dyes beyond their expiration dates for Mohs surgical patient testing in 2025 (5 of 5 months reviewed). The findings include: 1. Observation of the laboratory on 01/15/2026 at 09:00 a.m. revealed one grossing station used for gross examination of Mohs surgical tissue specimens. The station contained six different tissue-marking dyes. The following dyes had expired: Color Lot No Exp Date Blue 188140 12-31-2025 Green 176534 06-30-2025 Red 171349 04-30-2025 Orange 186726 11-30-2025 Black No lot or expiration indicated. 2. The laboratory director and Mohs technician confirmed the findings in an interview on 01/15/2026 at 11:00 a.m.</p>