

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D1026761	(X3) Date Survey Completed 08/13/2026
Name of Provider or Supplier Medical Care Pllc	Street Address, City, State 401 E Main Street, Johnson City, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The following deficiencies are a result of a desk review of proficiency testing scores obtained from the national database and verified with the proficiency testing company. The facility was found to be out of compliance with the conditions of the CLIA program. The following CONDITION LEVEL DEFICIENCIES were found to be out of compliance: D2016 - 42 C.F.R. 493.803 Condition: Successful participation [proficiency testing] D6000 - 42 C.F.R. 493.1403 Condition: Laboratories performing moderate complexity testing; laboratory director
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on a desk review of proficiency testing (PT) records from the Certification and

	Survey Provider Enhanced Reporting (CASPER) 0155 report and American Proficiency Institute (API) records for 2025 and 2026, the laboratory failed to successfully participate in a proficiency testing program approved by Health and Human Services (HHS), for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. The laboratory failed to successfully participate in the specialty of Chemistry for the Glucose and Total Iron analytes. Refer to D2096.
D2096	ROUTINE CHEMISTRY CFR(s): 493.841(f) (f) Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on a proficiency testing desk review of the CASPER 0155 report and the American Proficiency Institute (API) records for 2025 and 2026, the laboratory failed to achieve satisfactory performance (80% or better) for the same analyte in two out of three consecutive testing events for the Glucose analyte and in two consecutive testing events for the Total Iron analyte in the specialty of Chemistry. Findings included: 1. Review of the CASPER 0155 report revealed the following results: *Chemistry 2025-3rd Event: The laboratory received an unsatisfactory score of 0% for the Glucose analyte. *Chemistry 2026-1st Event: The laboratory received an unsatisfactory score of 60% for the Total Iron analyte. *Chemistry 2026 -2nd Event: The laboratory received an unsatisfactory score of 0% for the Glucose and Total Iron analytes. 2. A review of the API proficiency testing records confirmed the laboratory received the above results.
D6000	MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403 The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart. This CONDITION is not met as evidenced by: Based on a proficiency testing desk review of the CASPER 0155 report and the American Proficiency Institute (API) records for 2025 and 2026, the laboratory director failed to provide overall management and direction for laboratory services. The laboratory director failed to ensure that proficiency testing samples were tested as required. Refer to D6016.
D6016	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i) (e)(4)(i) The proficiency testing samples are tested as required under Subpart H of this part; This STANDARD is not met as evidenced by:

Based on a proficiency testing desk review of the CASPER 0155 report and the American Proficiency Institute (API) proficiency records for 2025 and 2026, the laboratory director failed to ensure that proficiency testing samples were tested as required. The laboratory director failed to ensure successful participation in an HHS-approved proficiency testing program. Refer to D2096.