

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D1056275	(X3) Date Survey Completed 04/03/2019
Name of Provider or Supplier Evans Medical Clinic, PLLC	Street Address, City, State 3493 Veteran'S Drive North Suite C, Huntingdon, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory procedure manual, final patient test reports and interview with the technical consultant, the laboratory failed to follow procedure for reporting wet prep in 2017, 2018, and 2019. The findings include: 1. Review of the laboratory procedure titled "WET PREPARATION" revealed the test report for wet prep should include trichomonas, yeast, and clue cells. 2. Review of patient numbers one (dated 12.15.17), two (dated 9.19.18), and three (dated 4.1.19) wet prep final patient reports revealed reporting for additional cells (white blood cells, red blood cells and epithelial cells) which were not included in the reporting protocol. 3. Interview with the technical consultant on April 3, 2019 at 2:00 pm confirmed the laboratory failed to follow the reporting protocol for wet prep in 2017, 2018, and 2019.</p>