

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D1071164	(X3) Date Survey Completed 11/19/2018
Name of Provider or Supplier Cardiovascular Clinic Of West Tn Pc	Street Address, City, State 2968 North Highland Ave, Jackson, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2089	<p>ROUTINE CHEMISTRY CFR(s): 493.841(c)</p> <p>Failure to participate in a testing event is unsatisfactory performance and results in a score of 0 for the testing event. Consideration may be given to those laboratories failing to participate in a testing event only if-- (1) Patient testing was suspended during the time frame allotted for testing and reporting proficiency testing results; (2) The laboratory notifies the inspecting agency and the proficiency testing program within the time frame for submitting proficiency testing results of the suspension of patient testing and the circumstances associated with failure to perform tests on proficiency testing samples; and (3)The laboratory participated in the previous two proficiency testing events.</p> <p>This STANDARD is not met as evidenced by: Based on review of the 2018 proficiency testing (PT) records, patient number one final report and interview with the technical consultant, the laboratory failed to participate in the 2018 proficiency testing events two and three for the lactate dehydrogenase (LDH). The findings include: 1) Review of the 2018 PT records revealed no participation for the LDH in events two and three. The 2018 event one documented corrective action stated the PT agency was contacted and informed no patient testing was performed. 2) Review of patient number one final report revealed the chemistry instrument printout, which includes the LDH, is scanned into the patient electronic medical record (EMR). 3) Interview on November 19, 2018 at 11:00 a.m. with the technical consultant confirmed the chemistry instrument printout with the LDH is scanned into the patient EMR.</p>