

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D1075485	(X3) Date Survey Completed 02/20/2020
Name of Provider or Supplier Skin Solutions Dermatology	Street Address, City, State 1040 N James Campbell Blvd Suite 101, Columbia, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5415	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(c)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies, as appropriate, must be labeled to indicate the following: (1) Identity and when significant, titer, strength or concentration. (2) Storage requirements. (3) Preparation and expiration dates. (4) Other pertinent information required for proper use.</p> <p>This STANDARD is not met as evidenced by: Based on observation of the laboratory and interview with the laboratory liaison, the laboratory failed to label dyes used for marking patient tissue samples with lot number and expiration date in 2020. The findings include: 1) Observation of the laboratory on February 20, 2020 at 8:45 a.m. revealed four dyes in use for marking patient tissue samples for MOHS surgery histopathology procedures. None of the four containers were labeled with lot numbers and expiration dates. 2) Interview with the laboratory liaison on February 20, 2020 at 11:15 a.m. confirmed the laboratory failed to label dye containers used for marking patient tissue samples with lot number and expiration date in 2020.</p>