Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	44D1089309	03/22/2019
Name of Provider or Supplier	**Street Address, City, State**	
Advanced Dermatology And Skin Cancer	7658 Poplar Pike, Germantown, TN	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6011	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(2) The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(2) and provide a safe environment in which employees are protected from physical, chemical, and biological hazards. This STANDARD is not met as evidenced by: Based on observation of the laboratory, review of the "Blood borne Pathogen Exposure Control Plan", and interview with processor number one, the laboratory director failed to ensure a safe environment for employees from biological hazards. The findings include: 1) Observation of the laboratory on March 22, 2019 at 10:26 a. m. revealed the laboratory area with four personal drinking containers in use on the counter. There were no biohazard safety signs and no 'No food or Drink Allowed' warning signs. 2) Review of the laboratory "Blood borne Pathogen Exposure Control Plan" revealed, "Training for the Exposure control Plan/Blood borne Pathogen must be provided for a new employee, an added/changed procedure and annually thereafter." 3) Interview on March 22, 2019 at 11:00 a.m. with processor number one confirmed the four containers of drinks on the laboratory counter next to the laptop computers. The "Blood borne Pathogen Exposure Control Plan" did not include a plan for biohazard safety. The employees did not have documented biohazard training.