

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D1103500	(X3) Date Survey Completed 08/21/2019
Name of Provider or Supplier Levy Dermatology, Pc	Street Address, City, State 6254 Poplar Avenue, Memphis, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation of the laboratory, review of patient number test report, and interview with the histotech, the laboratory failed ensure reagents were not used past their expiration date in 2019. The findings include: 1) Observation of the laboratory on August 21, 2019 at 8:45 am revealed red and blue dyes in use for marking patient tissue for histopathology procedures. The expiration date for the red dye was 09/18; the expiration date for the blue dye was 06/19. 2) Review of patient number six MOHS micrographic surgery report dated August 12, 2019 revealed tissue marking performed using both red and blue dyes. 3) Interview with the histotech on August 21, 2019 at 11:45 am confirmed the laboratory used the expired dyes used for marking patient tissue for histopathology procedures in 2019.</p>