

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D2032801	(X3) Date Survey Completed 12/11/2018
Name of Provider or Supplier Front Porch Pediatrics PLLC	Street Address, City, State 6847 Mountain View Rd, Ooltewah, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5403	<p>PROCEDURE MANUAL CFR(s): 493.1251(b)</p> <p>The procedure manual must include the following when applicable to the test procedure: (1) Requirements for patient preparation; specimen collection, labeling, storage, preservation, transportation, processing, and referral; and criteria for specimen acceptability and rejection as described in 493.1242. (2) Microscopic examination, including the detection of inadequately prepared slides. (3) Step-by-step performance of the procedure, including test calculations and interpretation of results. (4) Preparation of slides, solutions, calibrators, controls, reagents, stains, and other materials used in testing. (5) Calibration and calibration verification procedures. (6) The reportable range for test results for the test system as established or verified in 493.1253. (7) Control procedures. (8) Corrective action to take when calibration or control results fail to meet the laboratory's criteria for acceptability. (9) Limitations in the test methodology, including interfering substances. (10) Reference intervals (normal values). (11) Imminently life-threatening test results, or panic or alert values. (12) Pertinent literature references. (13) The laboratory's system for entering results in the patient record and reporting patient results including, when appropriate, the protocol for reporting imminently life threatening results, or panic, or alert values. (14) Description of the course of action to take if a test system becomes inoperable.</p> <p>This STANDARD is not met as evidenced by: ===== Based on review of the laboratory's procedure manual and upon interview with the Laboratory Director, determined the laboratory manual failed to include panic or alert values for the Complete Blood Count (CBC) test since 2016. The findings include: 1. A review of the laboratory's procedure manual revealed no panic or alert values included for CBC testing. 2. An interview at approximately 12:30 p.m. 12/11/18 with the Laboratory Director confirmed there were no panic or alert values included in the procedure manual for CBC testing. =====</p>