

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D2035674	(X3) Date Survey Completed 04/20/2022
Name of Provider or Supplier Pain Institute Of Nashville Plc	Street Address, City, State 1849 Madison St Suite D & Suite F, Clarksville, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6091	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(4)(iii)</p> <p>The laboratory director must ensure all proficiency testing reports received are reviewed by the appropriate staff to evaluate the laboratory's performance and to identify any problems that require corrective action.</p> <p>This STANDARD is not met as evidenced by: Based on observation of the laboratory, review of the laboratory's proficiency testing records and interview with the specimen processor, the laboratory director failed to ensure review of proficiency testing reports for 2021 event one and 2021 event three (two of three events). The findings include: 1. Observation of the laboratory on 04.20.2022 at 8 a.m. revealed patient testing for urine toxicology. 2. Review of the laboratory's proficiency testing records revealed the following: The laboratory is enrolled with College of American Pathologists for proficiency testing. 2021 event one had not been reviewed. 2021 event three had not been printed or reviewed. 3. Interview with the specimen processor on 04.20.2022 at 1:30 p.m. confirmed the laboratory director failed to ensure 2021 event one and 2021 event three proficiency testing results had been reviewed.</p>