

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D2065567	(X3) Date Survey Completed 11/26/2025
Name of Provider or Supplier Anycare 24	Street Address, City, State 702 South Cumberland St, Lebanon, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2009	<p>TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(1)</p> <p>(b)(1) The individual testing or examining the samples and the laboratory director must attest to the routine integration of the samples into the patient workload using the laboratory's routine methods.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the laboratory's American Proficiency Institute (API) proficiency testing (PT) records and staff interview, the testing personnel and the laboratory director/designee failed to sign one of eight attestation statements from 2023, 2024, and 2025. The findings include: 1. A review of eight of the laboratory's API PT testing events in 2023, 2024 and 2025 revealed that the testing personnel and the laboratory director/designee failed to sign an attestation statement for API 2023 Event 3. 2. The office manager and assistant office manager confirmed the survey findings in an interview on 11/25/2025 at 1:45 p.m. .</p>
D5211	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the laboratory's American Proficiency Institute (API) proficiency testing (PT) records and staff interviews, the laboratory failed to review and evaluate PT results for four of eight testing events in 2023, 2024, and 2025. The findings include: 1. A review of the laboratory ' s API PT records revealed no documented evaluation or review for the following events: - 2024 Hematology</p>

/Coagulation Event 2 - 2024 Hematology/Coagulation Event 3 - 2025 Hematology
/Coagulation Event 1 - 2025 Hematology/Coagulation Event 2 2. The office manager
and assistant office manager confirmed the survey findings during an interview on 11
/25/2025 at 1:45 p.m. .

D5417

TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT
CFR(s): 493.1252(d)

(d) Reagents, solutions, culture media, control materials, calibration materials, and
other supplies must not be used when they have exceeded their expiration date, have
deteriorated, or are of substandard quality.

This STANDARD is not met as evidenced by:

Based on laboratory observation, a review of quality control (QC) package inserts, QC
instrument printouts, and staff interviews, the laboratory failed to ensure that they did
not use hematology controls for complete blood count (CBC) testing past their
expiration date for two of five lot numbers that were in use in 2025. This deficiency is
a repeat deficiency identified in the previous survey conducted on 06/14/2023. The
findings include: 1. Observation of the laboratory revealed a Sysmex XP-300
hematology analyzer (Serial Number: A2587) and Eightcheck-3WP Xtra QC material
used for patient CBC testing. 2. A review of the QC package inserts revealed the
following lot numbers used in 2025. - Lot 4303 Expiration 02/05/2025 - Lot 5021
Expiration 04/30/2025 - Lot 5105 Expiration 07/23/2025 - Lot 5189 Expiration 10/15
/2025 - Lot 5273 Expiration 01/07/2026 3. A review of the Sysmex XP-300
instrument QC printouts revealed that the laboratory used the following QC lot
numbers past their expiration dates: - Lot 5021, Expiration 04/30/2025, used until 05
/07/2025, 7 days past expiration. - Lot 5189, Expiration 10/15/2025, used until 11/09
/2025, 24 days past expiration. 4. The office manager and assistant office manager
confirmed the survey findings during an interview on 11/25/2025 at 1:45 p.m.