

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D2120027	(X3) Date Survey Completed 05/30/2019
Name of Provider or Supplier Tennessee Oncology, Pllc	Street Address, City, State 4488 Carothers Parkway Suite 310, Franklin, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6046	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(8)</p> <p>(b) The technical consultant is responsible for-- (b)(8) Evaluating the competency of all testing personnel and assuring that the staff maintain their competency to perform test procedures and report test results promptly, accurately and proficiently.</p> <p>This STANDARD is not met as evidenced by: Based on review of personnel records and interview with the Technical Consultant (TC), the TC failed to perform competency assessments for the specialty Chemistry on 7 of 14 testing persons listed on the Centers for Medicare and Medicaid Services (CMS) 209 personnel form in 2017 and 2018. The findings include: 1. Review of personnel records revealed testing persons 3, 5, 6, 7, 8, 9 and 14 competency assessments were not performed by the TC for the specialty Chemistry in 2017 and 2018. 2. Interview with the TC on May 30, 2019 at 12:30 pm confirmed the TC failed to perform competency assessments for the specialty Chemistry on testing persons 3, 5, 6, 7,8, 9 and 14 in 2017 and 2018.</p>