

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D2141623	(X3) Date Survey Completed 05/20/2024
Name of Provider or Supplier Advanced Dermatology And Skin Cancer Assoc, PLLC	Street Address, City, State 5349 Airline Rd, Arlington, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory procedure manual, review of the Centers and Medicare and Medicaid Services Laboratory Personnel Report (CLIA) (Form CMS 209) and staff interview, the laboratory failed to have a policy/procedure for assessing testing personnel competency. The findings include: 1. The laboratory procedure manual did not have procedures for assessing testing personnel competency. 2. A review of Form CMS 209 revealed that one testing person, in addition to the laboratory director, interpreted histopathology slides. 3. The office manager confirmed the laboratory did not have procedures to assess testing personnel competency during an interview on 05/20/24 at 11:00 a.m.</p>