

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  44D2144906	<b>(X3) Date Survey Completed</b>  06/24/2025
<b>Name of Provider or Supplier</b>  Skin Solutions Dermatology	<b>Street Address, City, State</b>  800 Saundersville Rd, Hendersonville, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation of the laboratory, review of reagent logs, final patient test report, laboratory patient test logs, and staff interview, the laboratory failed to ensure that personnel did not use the Hematoxylin stain was not used beyond the manufacturer's expiration date for micrographically oriented histographic surgery (MOHS) histopathology patient testing for 19 days in 2024. The findings include: 1. Observation of the laboratory on 06/24/2025 at 9:30 a.m. revealed a Hematoxylin stain used to process patient tissue specimens for MOHS micrographic procedures. 2. A review of the laboratory's reagent log titled "Mohs Stain Expiration Log" for 2024, reviewed by the laboratory director on 12/30/2024, revealed the following: Hematoxylin lot number 2228732 was opened on 05/01/2023, expired on 01/25/2024, and was empty on 04/02/2024. Hematoxylin Lot number 2324408 was opened on 04/02/2024. The expired lot 2228732 was used for patient testing from 01/26/2024 through 04/01/2024. 3. A review of final patient test reports revealed that patient number 211264 was reported by the laboratory on 03/21/2024 when the expired Hematoxylin lot number 2228732 was used. 4. A review of the laboratory's patient log titled "Mohs Patient Daily Log" from January 2024 through April 2024 revealed the laboratory tested 89 patient samples from 01/26/2025 through 04/01/2025. 5. An interview with the practice manager on 06/24/2025 at 12:00 p.m. confirmed the survey findings.</p>