Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	44D2150401	10/07/2021
Name of Provider or Supplier	**Street Address, City, State**	
Redmed Of Arlington, Llc	5389 Airline Rd, Arlington, TN	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D6102	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(12) The laboratory director must ensure that prior to testing patients' specimens, all personnel have the appropriate education and experience, receive the appropriate training for the type and complexity of the services offered, and have demonstrated that they can perform all testing operations reliably to provide and report accurate results. This STANDARD is not met as evidenced by: Based on review of testing personnel records, patient test reports, and interview with the laboratory director, the laboratory director failed to ensure testing person number six received training and demonstrated accuracy prior to reporting patient test results in 2021. The findings include: 1. Review of testing personnel records revealed no documentation of training and demonstration of accuracy for testing person number six. 2. Review of patient test reports revealed seven patients were reported independently by testing person number six since employment began (patient chart numbers 120284, 94565, 142343, 201710, 198923, 198492, 81790). Dates tested spanned from 06.09.2021 to 09.08.2021. 3. Interview with the laboratory director on October 7, 2021 at 1:30 pm confirmed the laboratory director failed to ensure testing person number six received appropriate training and demonstrated accuracy prior to performing patient testing in 2021.