

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D2194690	(X3) Date Survey Completed 02/02/2022
Name of Provider or Supplier Octapharma Plasma, Inc	Street Address, City, State 7124 Winchester Rd, Memphis, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5211	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory records and staff interview, the laboratory failed to review total protein proficiency testing (PT) results for 2021 event two (one of three events reviewed). The findings include: 1. Review of the laboratory's proficiency testing records revealed the proficiency testing evaluation report for 2021 event two for total protein had not been reviewed (one of three PT events). 2. Interview on February 2, 2022 at approximately 9:30 am with the quality supervisor confirmed the laboratory failed to review proficiency testing results for total protein for 2021 event two.</p>