

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D2267035	(X3) Date Survey Completed 11/19/2025
Name of Provider or Supplier Skin Solutions Dermatology	Street Address, City, State 2130 Wilma Rudolph Blvd, Clarksville, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on laboratory observation, a review of laboratory records, and staff interview, the laboratory failed to ensure eosin stain used to process tissues removed during Mohs surgery was not used when the expiration date had been exceeded. The findings include: 1. Laboratory observation on 11/19/2025 at 10:00 a.m. revealed a Leica cryostat, inks, stains, and reagents used to process tissue removed during Mohs surgery in preparation for slide interpretation. During the observation, a container of eosin stain was noted that had expired (lot # 2323725, expiration date September 5, 2025). 2. A review of the facility's "H & E QA Control Log" revealed that approximately 48 Mohs surgery patients had slides stained during the period of expired stained use. 3. The Practice Administrator confirmed the findings during an interview on 11/19/2025 at 11:30 a.m.</p>