

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D2271068	(X3) Date Survey Completed 07/10/2023
Name of Provider or Supplier Tennessee Oncology Pllc	Street Address, City, State 4488 Carothers Pkwy Ste 120, Franklin, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5213	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(1)</p> <p>The laboratory must verify the accuracy of any analyte or subspecialty without analytes listed in subpart I of this part that is not evaluated or scored by a CMS-approved proficiency testing program.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's College of American Pathologists (CAP) proficiency testing (PT) records and staff interview, the laboratory failed to evaluate non-graded PT scores for minimal residual disease markers in 2022 (3 of 8 PT results). The findings include: 1. Review of the laboratory's CAP PT records for minimal residual disease (MRD-B 2022) revealed PT samples MRD1-04 and MRD1-06 had 3 ungraded results for the "BCR::ABL 1 p190 Log Red" tests. 2. Interview with the general supervisor on 07/10/23 at approximately 1:30 p.m. confirmed the laboratory failed to evaluate non-graded PT scores for accuracy in 2022 for 3 of the 8 minimal residual disease PT results reviewed.</p>