

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 44D2271068	(X3) Date Survey Completed 08/11/2025
Name of Provider or Supplier Tennessee Oncology Pllc	Street Address, City, State 4488 Carothers Pkwy Ste 120, Franklin, TN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6092	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(4)(iv)</p> <p>(e)(4)(iv) An approved corrective action plan is followed when any proficiency testing result is found to be unacceptable or unsatisfactory;</p> <p>This STANDARD is not met as evidenced by: Based on a review of the laboratory's College of American Pathologists (CAP) proficiency testing (PT) records, lack of records, and staff interviews, the laboratory director failed to document corrective action for unacceptable results received on one CAP PT event in 2024 (one of eight reviewed). The findings include: 1. A review of the laboratory's CAP PT records revealed that the laboratory received two unacceptable scores for the sample numbers NEO-04 and NEO-05 for the Neoplastic Cellularity (NEO) A 2024 event on the performance evaluation. 2. The laboratory did not have documented corrective action for the unacceptable results for the NEO-A 2024 samples NEO-04 and NEO-05 on the survey date (08/11/2025). 3. An interview with the laboratory general supervisor on 08/11/2025 at 2:45 p.m. confirmed the survey findings.</p>