

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  45D0495496	<b>(X3) Date Survey Completed</b>  01/28/2020
<b>Name of Provider or Supplier</b>  Cahrmc, Llc	<b>Street Address, City, State</b>  600 S Austin Road, Eagle Lake, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	The following deficiencies are a result of a desk review of proficiency testing scores obtained from the national database and verified with the proficiency testing company. The facility was found to be out of compliance with the conditions of participation of the CLIA program The following CONDITION LEVEL DEFICIENCIES were found to be out of compliance: 493.803 D2016 Condition: Successful participation [proficiency testing] 493.1403 D6000 Condition: Laboratories performing moderate complexity testing; laboratory director
<b>D2016</b>	<p><b>SUCCESSFUL PARTICIPATION</b> CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on a desk review of proficiency testing records obtained from the CMS (Center</p>

	<p>for Medicare Services) national database and verified with the proficiency testing company Wisconsin State Laboratory of Hygiene (WSLH), it was determined the laboratory had not successfully participated in a proficiency testing program approved by HHS, for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. The laboratory did not successfully participate in the specialty of Chemistry for the analyte Sodium. Refer to D2096.</p>
<p><b>D2087</b></p>	<p><b>ROUTINE CHEMISTRY</b> CFR(s): 493.841(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on a desk review of proficiency testing records obtained from the CMS (Center for Medicare Services) national database and verified with the proficiency testing records from Wisconsin State Laboratory of Hygiene (WSLH), it was determined the laboratory failed to attain a satisfactory score of at least 80% of acceptable responses for the analyte Sodium in 2019. Findings include: 1. WSLH PT 2019 Blood Gas - 2nd event the laboratory received an unsatisfactory score of 0% for Sodium. 2. WSLH PT 2019 Blood Gas - 3rd event the laboratory received an unsatisfactory score of 20% for Sodium.</p>
<p><b>D2096</b></p>	<p><b>ROUTINE CHEMISTRY</b> CFR(s): 493.841(f)</p> <p>Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on a desk review of proficiency testing records obtained from the CMS (Center for Medicare Services) national database and verified with the proficiency testing records from Wisconsin State Laboratory of Hygiene (WSLH), it was revealed the laboratory failed to achieve satisfactory performance for the analyte Sodium in two out of three consecutive testing events in 2019. Findings include: 1. WSLH PT 2019 Blood Gas - 2nd event the laboratory received an unsatisfactory score of 0% for Sodium. 2. WSLH PT 2019 Blood Gas - 3rd event the laboratory received an unsatisfactory score of 20% for Sodium. 3. Two out of three unsatisfactory scores results in unsuccessful proficiency testing performance.</p>
<p><b>D6000</b></p>	<p><b>MODERATE COMPLEXITY LABORATORY DIRECTOR</b> CFR(s): 493.1403</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on a desk review of proficiency testing records obtained from the CMS (Center</p>

for Medicare Services) national database and verified with the proficiency testing records from Wisconsin State Laboratory of Hygiene (WSLH), it was revealed that the laboratory director failed to provide overall management and direction of the laboratory services. Refer to D6018.

**D6018**

**LABORATORY DIRECTOR RESPONSIBILITIES**

CFR(s): 493.1407(e)(4)(iii)

The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(iii) Ensure that all proficiency testing reports received are reviewed by the appropriate staff to evaluate the laboratory's performance and to identify any problems that require corrective action;

This STANDARD is not met as evidenced by:

Based on a desk review of proficiency testing records obtained from the CMS (Center for Medicare Services) national database and verified with the proficiency testing records from Wisconsin State Laboratory of Hygiene (WSLH), it was revealed the laboratory director failed to ensure the overall quality of the laboratory services provided. The laboratory director failed to ensure successful participation in a HHS approved proficiency testing program. Refer to D2096.