

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  45D0503581	<b>(X3) Date Survey Completed</b>  01/12/2024
<b>Name of Provider or Supplier</b>  Sths Clinics Fm Edinburg Yarritu	<b>Street Address, City, State</b>  1200 S 10th Street, Edinburg, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	The following deficiencies are a result of a desk review of proficiency testing scores obtained from the CMS (Centers for Medicare and Medicaid Services) national database and verified with the proficiency testing company, American Proficiency Institute (API). The laboratory was found to be NOT in compliance with the conditions of participation of the CLIA program based on the following CONDITION LEVEL DEFICIENCIES: 493.803 Successful participation [proficiency testing] 493.1403 Laboratories performing moderate complexity testing; laboratory director
<b>D2016</b>	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on a desk review of proficiency testing records obtained from the CMS (Center</p>

	<p>for Medicare Services) national database and verified with the proficiency testing company, American Proficiency Institute (API), the laboratory had not successfully participated in a proficiency testing program approved by HHS, for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. The laboratory had unsuccessful participation in the specialty of hematology (refer to D2122).</p>
<p><b>D2122</b></p>	<p><b>HEMATOLOGY</b> CFR(s): 493.851(b)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on a desk review of the CMS 155 report and proficiency testing records from American Proficiency Institute (API), the laboratory failed to successfully participate for overall score in two consecutive testing events or two out of three consecutive testing events in the specialty of Hematology. The findings include: 1. A review of the CMS 155 report revealed the laboratory received the following unsatisfactory scores (passing = &gt;80%) for the specialty of hematology: Second testing event 2023 0% Third testing event 2023 0% 2. A desk review of the laboratory's American Proficiency Institute's results from the first event of 2023 and the third event of 2023 confirmed the proficiency testing scores: Second testing event 2023 0% Third testing event 2023 0%</p>
<p><b>D6000</b></p>	<p><b>MODERATE COMPLEXITY LABORATORY DIRECTOR</b> CFR(s): 493.1403</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on a desk review of laboratory proficiency testing performance, the laboratory director failed to provide overall management and direction of the laboratory services (refer to D6016).</p>
<p><b>D6016</b></p>	<p><b>LABORATORY DIRECTOR RESPONSIBILITIES</b> CFR(s): 493.1407(e)(4)(i)</p> <p>The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part;</p> <p>This STANDARD is not met as evidenced by: Based on a desk review of proficiency testing results, the laboratory director failed to ensure the overall quality of the laboratory services provided. The laboratory director</p>

failed to ensure successful participation in an HHS approved proficiency testing program (refer to D2122).