

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D0505253	(X3) Date Survey Completed 08/05/2021
Name of Provider or Supplier Adc, Pllc, Pediatrics At Cedar Bend, The	Street Address, City, State 2400 Cedar Bend, Austin, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D3011	<p>FACILITIES CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview, the laboratory failed to observe safety procedures by storing 9 cans of apple juice for patient consumption in 3 of 3 laboratory refrigerators. Findings follow. A. During a tour of the laboratory on August 5, 2021, the following refrigerators were found to contain apple juice: 1. Laboratory 1 refrigerator at 1010 hours 3 cans of apple juice stored in the same refrigerator as CBC (Complete Blood Count) controls, glucose reagents, and vaccines; 2. Laboratory 2 West refrigerator at 1020 hours 3 cans of apple juice stored in the same refrigerator as glucose reagents and vaccines; 3. Laboratory 2 East refrigerator at 1025 hours 3 cans of apple juice stored in the same refrigerator as CBC controls, glucose reagents and vaccines. B. Interview with the LVN Clinical Supervisor on August 5, 2021 at 1010 hours in the laboratory confirmed the apple juice was for the patients.</p>