

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D0678186	(X3) Date Survey Completed 05/18/2022
Name of Provider or Supplier Luis F Arango Md	Street Address, City, State 104 S Bryan, Mission, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The laboratory was surveyed on May 18, 2022 and found to be in compliance with the conditions of participation found in the CLIA regulations at 42 CFR 493 and recertification is recommended. The laboratory ceased testing moderate complexity chemistry testing. Therefore, the specialty will be closed.
D2121	<p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory policy, review of the laboratory's American Proficiency Institute (API) proficiency testing records, and confirmed in interview of laboratory personnel, the laboratory failed to attain a score of at least 80% for leukocyte count and white blood cell differential for one of four testing events reviewed. The finding were: 1. Review of the laboratory's policy titled 'Proficiency Testing' approved by the laboratory director on May 20, 2012 stated, "...Successful performance is considered to be passing 80% of all analytes in a subspecialty." 2. Review of the laboratory's API proficiency testing records found the laboratory received the following unsatisfactory analyte scores: 2022 (event 1) WBC Count 60% WBC Differential 33% 3. An interview with the technical consultant on May 18, 2022 at 14:00 hours in the break room confirmed the findings. Key: WBC - white blood cell</p>
D2122	<p>HEMATOLOGY CFR(s): 493.851(b)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p>

This STANDARD is not met as evidenced by:

Based on review of laboratory policy, review of the laboratory's American Proficiency Institute (API) proficiency testing records, and confirmed in interview of laboratory personnel, the laboratory failed to attain a score of at least 80% for overall testing score for one of four testing events reviewed. The findings were: 1. Review of the laboratory's policy titled 'Proficiency Testing' approved by the laboratory director on May 20, 2012 stated, "...Successful performance is considered to be passing 80% of all analytes in a subspecialty." 2. Review of the laboratory's API proficiency testing records found the laboratory received the following unsuccessful overall event score: 2022 (event 1) Hematology 78% 3. An interview with the technical consultant on May 18, 2022 at 14:00 hours in the break room confirmed the findings.