

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  45D0686795	<b>(X3) Date Survey Completed</b>  04/01/2025
<b>Name of Provider or Supplier</b>  Dermatology Consultants	<b>Street Address, City, State</b>  9 Medical Parkway Plaza 4 Suite # 105, Dallas, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	The laboratory was found to be in substantial compliance with CLIA regulations 42 CFR Part 493. Standard level deficiency cited.
<b>D5217</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor observation, review of twice annual accuracy assessments of Potassium Hydroxide (KOH) testing in 2023, and confirmed in interview, the laboratory failed to document two of two annual accuracy assessments of KOH testing in 2023. Findings Included: 1. During a tour of the facility on 04/01/2025 at 10:15 AM, the surveyor observed one bottle of KOH reagent available for patient testing in the microscope area. 2. Review of facility twice annual accuracy assessments of Potassium Hydroxide (KOH) testing in 2023, revealed no documentation of the laboratory performing two of two assessments in 2023. The surveyor requested the above documentation, and none was provided. 3. In an interview on 04/01/2025 at 09:48 AM in the facility breakroom, the laboratory histotechnician confirmed the above findings.</p>