

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 45D0688950	<b>(X3) Date Survey Completed</b> 11/16/2018
<b>Name of Provider or Supplier</b> National B Virus	<b>Street Address, City, State</b> 100 Piedmont Avenue, Atlanta, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on November 16, 2018. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiency was cited:
<b>D6046</b>	<p><b>TECHNICAL CONSULTANT RESPONSIBILITIES</b> CFR(s): 493.1413(b)(8)</p> <p>(b) The technical consultant is responsible for-- (b)(8) Evaluating the competency of all testing personnel and assuring that the staff maintain their competency to perform test procedures and report test results promptly, accurately and proficiently.</p> <p>This STANDARD is not met as evidenced by: Based on review of personnel competency assessment records and interviews with the technical Supervisor and laboratory director, the laboratory failed to include the six required competency assessment criteria when evaluating competency on testing personnel for Virology and Immunology testing procedures in 2017 and 2018. The findings include: 1. Review of testing personnel competency assessment records for 2017 and 2018 on (TP # 2 and 3 CMS 209), revealed the assessment did not include the six competency assessment criteria required by CLIA. 2. Interviews with the laboratory director and Technical Supervisor in the office used by the surveyor on November 16, 2018 at approximately 1:00 PM confirmed both employees annual competency assessments did not contain the six CLIA required competency assessment criteria.</p>