

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D0863781	(X3) Date Survey Completed 06/24/2026
Name of Provider or Supplier City Of Port Arthur Health Dept Lab	Street Address, City, State 5860 9th Ave, Port Arthur, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The laboratory was found to be in compliance with the Conditions of the CLIA regulations found at 42 CFR 493.1 through 493.1780, CLIA requirements for laboratories as a result of an recertification survey completed on June 24, 2026 and recertification is recommended. Standard level deficiencies were cited.
D6054	TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9) (b)(9) Thereafter, evaluations must be performed at least annually This STANDARD is not met as evidenced by: Based on review of laboratory personnel records, testing records, competency assessment records, and confirmed in interview, the technical consultant (TC) failed to ensure an annual competency assessment was performed for one of one testing personnel performing testing in 2025. The findings include: 1. Review of the laboratory personnel 209 form included one testing personnel (TP 1) performing the following patient testing: RPR, and RPR titer procedures for the screening of syphilis. Gram stain procedures, for the detection of intracellular coccobacilli. Wet Prep procedures, for the detection of trichomonas, yeast, and clue cells. 2. Review of laboratory competency assessments for 2025 did not include a competency assessment for TP 1 for patient testing performed. Surveyor asked for documentation the TC performed the 2025 annual competency assessment for RPR, RPR titers, Gram stains, and wet prep procedures, for TP1 and none could be provided. 3. In an interview on 6 /24/2026 at 13:20 hours, in the laboratory, the technical consultant confirmed 2025 competency assessment documentation was not available. KEY: RPR: Rapid Plasma Reagin Test