

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D0913782	(X3) Date Survey Completed 06/05/2026
Name of Provider or Supplier South Texas Dermatology	Street Address, City, State 4141 South Staples #300, Corpus Christi, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's twice annual accuracy records and confirmed in interview, the laboratory failed to verify the accuracy of diagnostic slide reading for two of two years (2024 and 2025). 1. Based on review of the laboratory's quality assurance records for twice annual accuracy, the laboratory failed to verify the accuracy of diagnostic slide reading in 2024 and 2025. Twice annual accuracy assessments were documented once in 2024 and once in 2025. 2. In a phone interview with the Pathology Coordinator on June 5, 2025 at 10:40 hours, she confirmed that the assessments were performed once in 2024 and once in 2025. .</p>