

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  45D0979623	<b>(X3) Date Survey Completed</b>  05/23/2019
<b>Name of Provider or Supplier</b>  Salas Minor Emergency Center	<b>Street Address, City, State</b>  1655 Ne Loop 286, Paris, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2089</b>	<p>ROUTINE CHEMISTRY CFR(s): 493.841(c)</p> <p>Failure to participate in a testing event is unsatisfactory performance and results in a score of 0 for the testing event. Consideration may be given to those laboratories failing to participate in a testing event only if-- (1) Patient testing was suspended during the time frame allotted for testing and reporting proficiency testing results; (2) The laboratory notifies the inspecting agency and the proficiency testing program within the time frame for submitting proficiency testing results of the suspension of patient testing and the circumstances associated with failure to perform tests on proficiency testing samples; and (3)The laboratory participated in the previous two proficiency testing events.</p> <p>This STANDARD is not met as evidenced by: Based on review of American Proficiency Institute (API) proficiency testing (PT) documentation for 2017, 2018 and 2019, confirmed by staff interview, the laboratory failed to participate in testing for routine chemistry analytes brain natriuretic peptide (BNP), creatinine kinase isoenzyme, D-dimer and troponin-I in the third event of 2018. Findings: 1. API PT documentation for 2018 showed scores of 0% in the third testing event for the following: BNP Creatinine Kinase, Isoenzyme D-Dimer Troponin-I Each score was noted "failure to participate" by API. 2. In an interview at the site on 05-23-2019, testing person 1 (CMS form 209) stated that she had forgotten to perform or assign performance of PT to other testing personnel for the third event 2018. She further stated that patient testing was not suspended during the time frame allotted for testing and reporting PT results.</p>