

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  45D0997959	<b>(X3) Date Survey Completed</b>  05/19/2026
<b>Name of Provider or Supplier</b>  Hill Country Dermatology	<b>Street Address, City, State</b>  2967 Oak Run Pkwy, Ste 210, New Braunfels, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor observation of reagents currently in use in the laboratory and staff interview, the laboratory failed to ensure expired materials were not used for testing. The findings included: 1. Surveyor observation of reagents in use in the laboratory on 05/19/2026 at 0930 hours identified two bottles of expired reagent: EDM3 Solutions Cholorazol Black E Lot: 3178 expiration date: 2025- 06-27 2. The MA confirmed the findings in an interview conducted on 05/19/2026 at 0945 hours in the laboratory.</p>