

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D1015161	(X3) Date Survey Completed 12/05/2019
Name of Provider or Supplier Tch Family Medical Clinic	Street Address, City, State 104 North Beech Street, Woodville, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: . Based on review of American Proficiency Institute (API) proficiency testing (PT) documentation for 2018 and 2019, confirmed by staff interview, the laboratory failed to successfully verify the accuracy of vaginal direct preparation slide interpretation in the second and third testing events of 2018. Findings: 1. Results for API PT for non-regulated tests in 2018 included the following: Microscopy: Vaginal Wet Preparation 2018 1st event: 100% 2018 2nd event: 0% 2018 3rd event: 0% Long Term Score: Unsuccessful 2. In an interview at the site on 12-05-2019, the laboratory director (CMS form 209) confirmed the reported results as accurate.</p>