Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D1032356	(X3) Date Survey Completed 05/02/2025
Name of Provider or Supplier All Children's Pediatric Clinic, Pa	**Street Address, City, State** 1313 S Closner Suite # A, Edinburg, TX	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	Based on a proficiency testing desk review survey performed on May 2, 2025, the laboratory was found to be out of compliance based on the following CONDITION LEVEL DEFICIENCIES: D2016 - 42 C.F.R. 493.803 Condition: Successful participation D6000 - 42 C.F.R. 493.1403 Condition: Laboratory Director, moderate complexity
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on review of the Certification and Survey Provider Enhanced Reporting (CASPER) Report 155 Individual Laboratory Profile and American Proficiency Institute's proficiency reports, the laboratory failed to achieve satisfactory

	performance in two of three testing events for the specialty of hematology (refer to D2131) and the analyte of hemoglobin (refer to D2130) resulting in an initial unsuccessful performance.
D2130	HEMATOLOGY CFR(s): 493.851(f) (f) Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on review of the Certification and Survey Provider Enhanced Reporting (CASPER) Report 155 Individual Laboratory Profile and American Proficiency Institute's proficiency reports, the laboratory failed to achieve satisfactory performance for two of three events in 2024 and 2025 for the analyte of hemoglobin. The findings included: 1. Based on review of the Certification and Survey Provider Enhanced Reporting (CASPER) Report 155 Individual Laboratory Profile report, the laboratory received the following unsatisfactory performances for hemoglobin in two of three events: 2024 API 2nd event 60% 2025 API 1st event 20% 2. Based on review of the American Proficiency Institute's proficiency reports, the laboratory received the following unsatisfactory performances for hemoglobin in two of three events: 2024 API 2nd event 60% 2025 API 1st event 20%
D2131	HEMATOLOGY CFR(s): 493.851(g) (g) Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on review of the Certification and Survey Provider Enhanced Reporting (CASPER) Report 155 Individual Laboratory Profile and American Proficiency Institute's proficiency reports, the laboratory failed to achieve satisfactory performance for two of three events in 2024 and 2025 for the specialty hematology. The findings included: 1. Based on review of the Certification and Survey Provider Enhanced Reporting (CASPER) Report 155 Individual Laboratory Profile report, the laboratory received the following unsatisfactory performances for the specialty of hematology in two of three events: 2024 API 2nd event 78% 2025 API 1st event 45% 2. Based on review of the American Proficiency Institute's proficiency reports, the laboratory received the following unsatisfactory performances for specialty hematology in two of three events: 2024 API 2nd event 78% 2025 API 1st event 45%
D6000	MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403 The laboratory must have a director who meets the qualification requirements of 493. 1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.

	This CONDITION is not met as evidenced by: Based on desk review of the Certification and Survey Provider Enhanced Reporting (CASPER) Report 155 Individual Laboratory Profile and American Proficiency Institute's proficiency reports, the laboratory director failed to provide overall management and direction of the laboratory services. Refer to D6016.
D6016	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i) (e)(4)(i) The proficiency testing samples are tested as required under Subpart H of this part; This STANDARD is not met as evidenced by: Based on desk review of the Certification and Survey Provider Enhanced Reporting (CASPER) Report 155 Individual Laboratory Profile and American Proficiency Institute's proficiency reports, the laboratory director failed to ensure successful participation in a HHS approved proficiency testing program for the specialty hematology (refer to D2131) and the analyte of hemoglobin (refer to D2130) for two of three events in 2024 and 2025, resulting in an initial unsuccessful performance.