Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 45D1061372	**(X3) Date Survey Completed** 07/13/2021
Name of Provider or Supplier Center For Cancer & Blood Disorders -	**Street Address, City, State** 7415 Las Colinas Blvd Suite 100, Irving, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D0000	Laboratory representatives were present at the entrance conference. The survey process was discussed. An opportunity for questions and comments was given. The exit conference was held with the laboratory representatives. The laboratory was found to be in substantial compliance for the specialties/subspecialties for which it was surveyed. The standard level deficiencies cited were discussed. The process for submitting the corrections was explained. CMS form 2567 will be emailed from the Texas Health and Human Services Commission, Health Facility Compliance Arlington Group. Note: The CMS-2567 (Statement of Deficiencies) is an official, legal document. All information must remain unchanged except for entering the plan of correction, correction dates, and the signature space. Any discrepancy in the original deficiency citation(s) will be reported to the Dallas Regional Office (RO) for referral to the Office of the Inspector General (OIG) for possible fraud. If information is inadvertently changed by the provider/supplier, the State Survey Agency (SA) should be notified immediately.
D6053	TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9) The technical consultant is responsible for evaluating and documenting the performance of individuals responsible for moderate complexity testing at least semiannually during the first year the individual tests patient specimens. This STANDARD is not met as evidenced by: Based on review of laboratory policy, personnel records and confirmed in interview, the technical consultant failed to evaluate and document the performance for 1 of 4 Testing Persons (TP-1) responsible for moderate complexity testing at least semiannually during the first year the individual tests patient specimens in 2019. Findings: 1. Review of the laboratory's policy titled "Personnel Training and Competency" revealed: "Annual Evaluation and Competency Each employee will be given an evaluation for competency after training, at six months from initial

	competency and annually thereafter." 2. Review of personnel records for TP-1 revealed an initial training assessment was performed on 05/09/2019 and an annual competency assessment performed on 06/30/2020. There was no documentation of a semiannual performance (due 11/2019). The technical consultant did not ensure semi-annual competency assessments were evaluated and documented for TP-1. 3. During an interview on 07/13/2021 at 9:24 am, the Technical Consultant confirmed the above findings.