

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 45D1063733	(X3) Date Survey Completed 03/30/2022
Name of Provider or Supplier First Surgical Hospital	Street Address, City, State 4801 Bissonnet Street, Bellaire, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The following deficiencies are a result of a desk review of proficiency testing scores obtained from the national database and verified with the proficiency testing company. The facility was found to be out of compliance with the conditions of participation of the CLIA program The following CONDITION LEVEL DEFICIENCIES were found to be out of compliance: 493.803 D2016 Condition: Successful participation [proficiency testing] 493.807 D2017 Reinstatement After Failure 493.1403 D6000 Condition: Laboratories performing moderate complexity testing; laboratory director
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by:</p>

Based on a desk review of proficiency testing records obtained from the CMS (Center for Medicare Services) national database and verified with the proficiency testing company American Proficiency Institute (API) and the facility, it was determined the laboratory had not successfully participated in a proficiency testing program approved by CMS, for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. The laboratory did not successfully participate in the specialty of Routine Chemistry for the analyte Chloride. (Refer to D2096)

D2017

REINSTATEMENT OF NONWAIVED LABORATORIES
CFR(s): 493.807(a)(b)

(a) If a laboratory's certificate is suspended or limited or its Medicare or Medicaid approval is cancelled or its Medicare or Medicaid payments are suspended because it fails to participate successfully in proficiency testing for one or more specialties, subspecialties, analyte or test, or voluntarily withdraws its certification under CLIA for the failed specialty, subspecialty, or analyte, the laboratory must then demonstrate sustained satisfactory performance on two consecutive proficiency testing events, one of which may be on site, before CMS will consider it for reinstatement for certification and Medicare or Medicaid approval in that specialty, subspecialty, analyte or test. (b) The cancellation period for Medicare and Medicaid approval or period for suspension of Medicare or Medicaid payments or suspension or limitation of certification under CLIA for the failed specialty, subspecialty, or analyte or test is for a period of not less than six months from the date of cancellation, limitation or suspension of the CLIA certificate.

This CONDITION is not met as evidenced by:

Based on a desk review of proficiency testing records obtained from the CMS (Center for Medicare Services) national database (Report 153 and Report 155), and verified with the proficiency testing records from American Proficiency Institute (API) and the facility, it was determined the laboratory had not successfully participated in proficiency testing for the analyte Chloride under the specialty of Routine Chemistry for 4 of 6 testing events and has not demonstrated sustained satisfactory performance on two consecutive proficiency events since the unsuccessful scores. Findings included: 1. 2020 Chemistry - Core - second event the laboratory received an unsatisfactory score of 0% for the analyte Chloride. 2. 2020 Chemistry - Core - third event the laboratory received an unsatisfactory score of 60% for the analyte Chloride. * The 2020 second event and 2020 third event resulted in initial unsuccessful proficiency testing performance. 3. 2021 Chemistry - Core - second event the laboratory received an unsatisfactory score of 20% for the analyte Chloride. * The 2021 second event constituted non-initial unsuccessful proficiency testing performance. 4. 2022 Chemistry - Core - first event the laboratory received an unsatisfactory score of 60% for the analyte Chloride. * The 2022 first event constituted an additional non-initial unsuccessful proficiency testing performance. Note: The laboratory has voluntarily ceased testing Chloride on their primary instrument, The Horiba Pentra C400 on 02/23/2022.

D2096

ROUTINE CHEMISTRY
CFR(s): 493.841(f)

Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

	<p>This STANDARD is not met as evidenced by: Based on a desk review of proficiency testing records obtained from the CMS (Center for Medicare Services) national database Report 153 and Report 155, and verified with the proficiency testing records from American Proficiency Institute (API), it was revealed the laboratory failed to achieve satisfactory performance for the analyte Chloride for 4 of 6 testing events in 2020, 2021 and 2022. Findings included: 1. 2020 Chemistry - Core - second event the laboratory received an unsatisfactory score of 0% for the analyte Chloride. 2. 2020 Chemistry - Core - third event the laboratory received an unsatisfactory score of 60% for the analyte Chloride. 3. 2021 Chemistry - Core - second event the laboratory received an unsatisfactory score of 20% for the analyte Chloride. 4. 2022 Chemistry - Core - first event the laboratory received an unsatisfactory score of 60% for the analyte Chloride. Four out of six unsatisfactory scores results in non- initial unsuccessful proficiency testing performance.</p>
<p>D6000</p>	<p>MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on a desk review of proficiency testing records obtained from the CMS (Center for Medicare Services) national database and verified with the laboratory's American Proficiency Institute (API) proficiency test results, it was revealed that the laboratory director failed to provide overall management and direction of the laboratory services. (Refer to D6016)</p>
<p>D6016</p>	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i)</p> <p>The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part;</p> <p>This STANDARD is not met as evidenced by: Based on a desk review of proficiency testing records obtained from the CMS (Center for Medicare Services) national database and verified with the laboratory's American Proficiency Institute (API) proficiency test results, it was revealed the laboratory director failed to ensure the overall quality of the laboratory services provided. The laboratory director failed to ensure successful participation in a CMS approved proficiency testing program. (Refer to D2096)</p>